

# Blind Source Separation of Neuromagnetic Responses

Akaysha C. Tang <sup>a,b,c</sup> Barak A. Pearlmutter <sup>b,c</sup>  
Michael Zibulevsky <sup>b</sup>

<sup>a</sup>*Dept of Psychology*, <sup>b</sup>*Dept of Computer Science*, <sup>c</sup>*Dept of Neurosciences*,  
*University of New Mexico, Albuquerque, NM 87131*

---

## Abstract

Magnetoencephalography (MEG) is a functional brain imaging technique with millisecond temporal resolution and millimeter spatial resolution. The high temporal resolution of MEG compared to fMRI and PET (milliseconds vs. seconds and tens of seconds) makes it ideal for measuring the precise time of neuronal responses, thereby offering a powerful tool for studying temporal dynamics. We applied blind source separation (BSS) to continuous 122-channel human magnetoencephalographic data from two subjects and five tasks. We demonstrate that without using any domain specific knowledge and without making the common assumption of single or multiple current dipole sources, BSS is capable of separating non-neuronal noise sources from neuronal responses and also of separating neuronal responses from different sensory modalities, and from different processing stages within a given modality.

*Key words:* functional brain imaging; ICA; MEG; blind source separation.

---

## 1 Introduction

The brain's neuromagnetic signals are small compared to the noise (signal strength: 100fT; sensor noise: 10fT; noise in a shielded environment from magnetized lung contaminants: 10<sup>6</sup>fT; from abdominal currents: 10<sup>5</sup>fT; from heart and eyeblinks: 10<sup>4</sup>fT; from epileptic and spontaneous activity: 10<sup>3</sup>fT (8)) and therefore neuronal responses evoked in cognitive tasks are usually averaged across many trials of stimulation. Recently, a number of blind source separation (BSS) algorithms have successfully separated mixtures in other domains (5, 9) without any special knowledge of the nature of the mixed signals or the mixing process. Most relevant to multichannel brain imaging data, Bell-Sejnowski Infomax (1995) and Fast ICA (6) have successfully isolated and removed major artifacts from EEG and MEG data (7, 11).

In MEG, magnetic activity from different neuronal populations is observed by sensors arranged around the subject’s head. Each sensor responds to a linear mixture of the signals emitted by multiple sources. Responses from different neuronal populations will be referred to as *source responses* and the neuronal populations that give rise to these responses will be referred to as *neuronal sources* or simply *sources*. These neuronal sources often, but not necessarily, consist of a spatially contiguous population of neurons. BSS separates the measured sensor signals into components which are as independent as possible, each with its own field map. Using spatial and temporal constraints, we will establish a correspondence between the BSS components and noise sources, and between BSS components and various neuronal sources.

## 2 Methods

Two volunteer subjects performed a total of five visual reaction time tasks which differed in the amount of processing required for target determination. In all but one task, extraction of both object identity and object location was required. Each trial consisted of a pair of colored abstract block compositions, presented symmetrically and simultaneously on the left and right halves of a screen. Subjects were instructed to respond as quickly and accurately as possible with a left- or right-hand mouse button press when a target stimulus was presented to the left or right side of the screen, respectively.

Neuromagnetic signals were recorded continuously throughout the experiment at 300Hz using a whole-head 122-channel direct-current superconducting quantum interference device (DC-SQUID) neuromagnetometer (Neuromag, Helsinki, Finland.) A band-pass filter (0.1–100Hz) was applied. This high dimensionality and large volume of data of continuous MEG data imposes a severe computational burden. After experimenting with the Infomax, cICA and Fast ICA algorithms, we settled on a variant of the SOBI algorithm (2), which gave sufficient accuracy (in the sense of properly separating out known and previously unknown sources of noise) at a reasonable computational cost.

## 3 Results

The BSS process separates out as many components as there are sensors, in this case 122 components. Some of these correspond to responses of neuronal populations that are either task-dependent or task-independent; others, to various noise sources such as muscle activity due to eyeblinks, 60Hz from electrical equipment, slow drift due to unknown causes, signals from dental fillings, sensor quantum jumps, and other unknown noise sources associated

with random neuronal activity. Many recovered components are due to random factors not under experimental control and can be discarded. Only a subset of the separated components can be identified using their spatial and temporal properties and/or prior knowledge of the experiment.

BSS components are characterized by both *MEG images* and *field maps*. The MEG image provides temporal information and the field map provides spatial information about a particular component. When both the temporal and spatial information are consistent with anatomical and neurophysiological properties of a known neuronal population, the activity of the component is considered to reflect the activity of the known neuronal population. Confirmation and verification of component responses associated with different processing stages, *e.g.* early and late visual components, are provided by satisfying the temporal and spatial constraints on early and later visual processing set by our knowledge of the visual system.

Figure 1 shows examples of known noise sources (**a**, **b**), eyeblinks (**c**), visual sources (**d**, **e**), and somatosensory sources (**f**, **g**). The 60Hz source contained exactly 60 cycles in a window of 1000ms, and had a rather distributed field map. The slow DC drifting source did not show any stimulus-locked changes but changed slowly across trials, and also had a rather distributed field map. Eyeblinks were confirmed by an independently collected EOG signal and by the field activation around the relevant muscle groups. Visual sources associated with early and later processing were established by both their response latency (see MEG images) and the occipital and occipital-parietal activation of their field maps. Somatosensory sources (7, 34) were identified by their button-press locked responses, and are shown aligned by button press (1000ms pre- and post-button press). Their field maps are consistent with activation of the right and left somatosensory cortex. These typical noise and neuronal sources were consistently identified in all five experiments performed by both subjects.

## 4 Discussion

We applied a blind source separation algorithm to continuous 122-channel MEG data. The separated components include identifiable known noise sources and neurophysiologically and neuroanatomically meaningful neuronal components. These neuronal components can be viewed as “cleaned” source responses, and can be fed to any existing source localization algorithm for further localization using co-registration of MEG and MRI. Since these neuronal components are already separated from many noise sources and from each other, we expect improved accuracy in future source localization, when blind source separation is coupled with well-developed source localization methods.




Fig. 1. Separation of noise sources and neuronal components using blind source separation. Left: *MEG images*. Each panel contains 90 trials. Trials are ordered vertically, except for (c), which is sorted by reaction time and where the curve on the right indicates the times of button presses. Pixels are color-coded source strength. Vertical bars on the left indicate stimulus onset time. Pre- and post-stimulus durations are each 1000ms. Right: *field maps*. Left sagittal, right sagittal, dorsal, and posterior views. False colors indicate the strength with which the component influences the 61 sensor pairs. MEG images and field maps of (a) a 60Hz component, (b) a slow DC drift component, (c) an eyeblink component, (d) an early occipital visual component (27) (e) a later occipital-parietal visual component (14) (f) a right somatosensory component (7) (g) a left somatosensory component (34).

One fundamental difference between BSS and source localization algorithms traditionally used for MEG (4) is that BSS does not assume any prior physical source model. BSS algorithms only require temporal information from the sensor signals to generate an approximation of the original source signals, without making reference to or conclusions about spatial locations of the sources. As a result, BSS can deal equally well with source signals generated by a population of neurons that are best described as a point, a cylinder, a large wrinkled sheet, or neuronal populations discretely distributed. The fact that BSS does not require a physical model for its sources guarantees that there will be no separation errors arising from false or non-optimal physical source models. This flexibility also permits the use of a wider and more ecological range of stimuli, which often do not generate well-defined focal activation.

The ability of the BSS algorithm to separate out various noise sources and physiologically and anatomically interpretable neuronal components arises from its use of the temporal characteristics present in continuous MEG data. SOBI, the algorithm used in this study, assumes that the components are independent across multiple time scales and tries to minimize the temporal correlation at these time scales. Of course, the assumptions made in the derivation of the blind source separation algorithm—namely statistical independence—is not 100% met in the MEG domain. However, there are theoretical reasons to believe that BSS algorithms of this sort are quite robust (3), *i.e.* the goodness of the separation suffers little when assumptions are not completely met.

Therefore it is reasonable to test whether these algorithms are actually capable of successfully separating this data in practice, just as t-tests are, in practice, used on non-Gaussian data due to their robustness. The physiological interpretability of our separated components is consistent with the notion that BSS algorithms are sufficiently robust in practice to obtain good performance on MEG data. Given that independent sources of noise exist within each individual neuron, and that there are regional differences in cell types, packing density, transmitter receptor distribution, and feed-forward/feed-back/lateral connectivity (10, 12), it is not surprising that BSS is sufficiently robust to yield physiologically and anatomically meaningful results.

### *Acknowledgments*

This research was supported in part by NSF CAREER award 97-02-311, the National Foundation for Functional Brain Imaging, an equipment grant from Intel corporation, the Albuquerque High Performance Computing Center, a gift from George Cowan, and a gift from the NEC Research Institute. We thank Scott Carter, Dan Phung, Greg Kirk, and Brian Sanchez for programming support, Michael Weisend, Robert Christner, Jeffrey Meyer, and Marianna Lanoue for their assistance in data collection.

## References

- [1] AJ Bell and TJ Sejnowski. An information-maximization approach to blind separation and blind deconvolution. *Neural Computation*, 7(6): 1129–1159, 1995.
- [2] A Belouchrani, KA Meraim, J-F Cardoso, and E Moulines. Second-order blind separation of correlated sources. In *Proc. Int. Conf. on Digital Sig. Proc.*, pages 346–351. Cyprus, 1993.
- [3] J-F Cardoso. Blind signal separation: statistical principles. *Proceedings of the IEEE*, 9(10):2009–2025, October 1998.
- [4] M Hamalainen, R Hari, RJ Ilmoniemi, J Knuutila, and OV Lounasmaa. Magnetoencephalography—theory, instrumentation, and applications to noninvasive studies of the working brain. *Rev. Modern Physics*, 65:413–497, 1993.
- [5] A Hyvarinen. Survey on independent component analysis. *Neural Computing Surveys*, 2:94–128, 1999.
- [6] A Hyvarinen and E Oja. A fast fixed-point algorithm for independent component analysis. *Neural Computation*, 9(7), October 1997.
- [7] T-P Jung, C Humphries, T-W Lee, MJ McKeown, V Iragui, S Makeig, and TJ Sejnowski. Removing electroencephalographic artifacts by blind source separation. *Psychophysiology*, 1999. In Press.
- [8] JD Lewine and WW Orrison, II. Magnetoencephalography and magnetic source imaging. In WW Orrison, II, JD Lewine, JA Sanders, and MF Hartshorne, editors, *Functional Brain Imaging*, pages 369–417. Mosby, St. Louis, 1995.
- [9] BA Pearlmutter and LC Parra. A context-sensitive generalization of ICA. In *International Conference on Neural Information Processing*, pages 151–157, Hong Kong, September 24–27 1996. Springer-Verlag.
- [10] KS Rockland. Elements of cortical architecture. In KS Rockland, JH Kaas, and A Peters, editors, *Cerebral Cortex. Volume 12. Extrastriate cortex in primates*, pages 243–293. Plenum Press, New York, 1997.
- [11] R Vigário, V Jousmäki, M Hämäläinen, R Hari, and E Oja. Independent component analysis for identification of artifacts in magnetoencephalographic recordings. In *Advances in Neural Information Processing Systems 10*. MIT Press, 1998.
- [12] K Zilles and S Clarke. Architecture, connectivity, and transmitter receptors of human extrastriate visual cortex. In KS Rockland, JH Kaas, and A Peters, editors, *Cerebral Cortex. Volume 12. Extrastriate cortex in primates*, pages 673–742. Plenum Press, New York, 1997.